

## WaterMark Certification Scheme

### Risk Assessment Notification

**2017/R35**

**Product Assessed: - Plate Heat Exchanger**

Level 12, 224 Bunda Street  
Canberra ACT 2600  
GPO Box 2013  
CANBERRA ACT 2601 Australia  
Phone: 1300 134 631  
Email: [NCC@abcb.gov.au](mailto:NCC@abcb.gov.au)  
Web: [www.abcb.gov.au](http://www.abcb.gov.au)

### Assessment

The WaterMark Administration advises it has reviewed a risk assessment submitted for the above mentioned product. The product is utilised for transferring heat from a primary source to a secondary supply notably drinking water applications by utilising plates interlocked and stacked to provide maximum heat transfer area.

The WaterMark Administration verifies that the risk assessment for the plate heat exchanger has identified risks requiring mitigation through the WaterMark Certification Scheme.

### Outcome

The Risk Assessment identified the overall risk rating being medium for the nominated health and safety consequences incidences. In accordance with the *Manual for the WaterMark Certification Scheme, Appendix 3, Table D*, any product with an overall risk rating of medium requires WaterMark Certification.

Approved WaterMark Conformity Assessment Bodies will now need to evaluate these products to the most applicable specification. If the identified risks are unable to be satisfactorily addressed through an existing Specification, then a new, or amended Specification shall be proposed. A submission to the ABCB shall be undertaken in accordance with the *Manual for the WaterMark Certification Scheme, Appendix 4, Protocol for Developing Product Specifications*.

### Issued

14 November 2018.

- END -